
Midmark M11 - v1.0.0
Total Cycles: 2476

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2022 10:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.6 F 0.0
2:00 191.9 F 3.8
4:00 197.8 F 8.4
6:00 221.8 F 13.4
8:00 227.1 F 7.4
10:00 239.3 F 12.6
12:00 237.1 F 9.8
14:00 238.7 F 8.4
16:00 249.4 F 14.9
18:00 261.5 F 21.7
20:00 272.5 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.4 F 29.0
2:00 271.5 F 28.2
3:00 272.5 F 29.1
4:00 271.9 F 28.5

Min 271.3 F 28.2
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 222.6 F 0.0
10:00 215.2 F 0.0
15:00 233.5 F 0.0
20:00 237.5 F 0.0
25:00 227.8 F 0.0
30:00 217.7 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:56
CYCLE COMPLETE

10/24/2022 11:05 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____