
Midmark M11 - v1.0.0
Total Cycles: 1898

Name: RITTER 1
Sterilizer ID: V2178198
05/03/2022 07:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 72.6 F 0.0
4:00 82.0 F 0.0
6:00 115.4 F 0.1
8:00 146.5 F 1.0
10:00 171.1 F 3.1
12:00 187.8 F 5.8
14:00 203.7 F 8.9
16:00 218.2 F 12.5
18:00 224.2 F 5.7
20:00 233.7 F 9.8
22:00 239.6 F 9.9
24:00 241.8 F 11.3
26:00 237.9 F 9.1
28:00 251.2 F 15.6
30:00 262.1 F 21.8
32:00 272.0 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.0
1:00 272.6 F 28.7
2:00 271.5 F 27.9
3:00 272.3 F 28.8
4:00 272.1 F 28.2

Min 271.2 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.0 F 0.0
5:00 216.8 F 0.0
10:00 217.2 F 0.0
15:00 205.6 F 0.0
20:00 213.8 F 0.0
25:00 211.5 F 0.0
30:00 214.6 F 0.0

Drying Complete

FILLING: 0:46
HEATING: 32:34
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:10:27
CYCLE COMPLETE

05/03/2022 08:35 AM
ACCEPTED: X
REJECTED:
Operator: U02