
Midmark M11 - v1.0.5
Total Cycles: 102

Name: RITTER2
Sterilizer ID: V2313228
06/15/2021 12:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.0 F 0.0
2:00 127.6 F 1.5
4:00 147.0 F 4.2
6:00 174.6 F 8.2
8:00 197.0 F 12.9
10:00 213.0 F 5.9
12:00 226.7 F 10.6
14:00 230.7 F 7.8
16:00 242.0 F 13.3
18:00 242.6 F 12.3
20:00 254.9 F 18.6
22:00 265.9 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.4 F 29.8
2:00 271.6 F 29.2
3:00 272.7 F 29.9
4:00 272.0 F 29.3

Min 271.5 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.2 F 0.0
5:00 219.2 F 0.0
10:00 223.0 F 0.0
15:00 225.4 F 0.0
20:00 225.0 F 0.0
25:00 224.4 F 0.0
30:00 223.6 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 23:44
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:02:12
CYCLE COMPLETE

06/15/2021 01:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____