

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3174

Name: RITTER 1  
Sterilizer ID: V2178198  
04/28/2023 08:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
2:00 186.6 F 0.6  
4:00 197.7 F 4.3  
6:00 216.2 F 9.7  
8:00 225.0 F 7.1  
10:00 235.6 F 10.5  
12:00 234.1 F 7.6  
14:00 242.7 F 11.4  
16:00 246.7 F 13.4  
18:00 259.5 F 20.4  
20:00 270.9 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.1  
1:00 272.4 F 28.5  
2:00 272.1 F 28.6  
3:00 272.2 F 28.4  
4:00 272.1 F 28.7  
  
Min 271.5 F 28.0  
Max 272.8 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 228.0 F 0.0  
10:00 227.1 F 0.0  
15:00 225.8 F 0.0  
20:00 224.6 F 0.0  
25:00 225.6 F 0.0  
30:00 224.9 F 0.0  
Drying Complete

FILLING: 1:44  
HEATING: 20:42  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 00:59:42  
CYCLE COMPLETE

04/28/2023 09:55 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_