

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3059

Name: RITTER 1  
Sterilizer ID: V2178198  
03/30/2023 12:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.9 F 0.0  
2:00 198.5 F 4.9  
4:00 205.4 F 10.2  
6:00 221.4 F 5.8  
8:00 233.1 F 11.2  
10:00 234.6 F 8.5  
12:00 239.8 F 9.9  
14:00 247.9 F 14.2  
16:00 260.7 F 21.2  
18:00 271.9 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.3  
1:00 272.7 F 29.1  
2:00 271.9 F 28.5  
3:00 271.5 F 28.5  
4:00 272.6 F 29.1  
  
Min 271.3 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.3 F 0.0  
5:00 227.0 F 0.0  
10:00 242.7 F 0.0  
15:00 239.0 F 0.0  
20:00 228.7 F 0.0  
25:00 222.2 F 0.0  
30:00 220.7 F 0.0  
Drying Complete

FILLING: 2:04  
HEATING: 18:35  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:57:57  
CYCLE COMPLETE

03/30/2023 12:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_