

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
2.3uS / 1.4ppm  
CYCLE NUMBER 003679  
9:18 25/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:31  
PRESSURIZING 2:34  
STERILIZING 3:36  
136.4 C 223kPa 3:36  
136.3 C 224kPa 4:06  
136.3 C 223kPa 4:36  
136.4 C 224kPa 5:06  
136.4 C 222kPa 5:36  
136.4 C 223kPa 6:06  
136.4 C 223kPa 6:36  
136.0 C 224kPa 7:06  
136.0 C 224kPa 7:06  
VENTING 7:06  
AIR DRYING 7:55  
CYCLE COMPLETE 9:40

-----  
Digital Signature #  
BC5A53BB6CD9AA0C  
-----