

-----  
Surrey, BC  
Canada  
-----

STATIM 5000 S501R711  
SN 511018F00076  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000051  
13:11 21/09/2020  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 6:04  
PRESSURIZING 8:25  
STERILIZING 10:30  
136.0 C 222kPa 10:30  
136.4 C 226kPa 11:01  
136.5 C 227kPa 11:31  
136.4 C 227kPa 12:01  
136.3 C 225kPa 12:31  
136.4 C 226kPa 13:01  
136.3 C 225kPa 13:31  
136.3 C 225kPa 14:00  
VENTING 14:01  
AIR DRYING 15:15  
CYCLE COMPLETE 45:16  
-----

Digital Signature #  
C138F55D39FDBCDC  
-----