
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005610
9:03 29/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:39
PRESSURIZING 2:41
STERILIZING 3:38
136.4 C 222kPa 3:38
136.4 C 225kPa 4:09
136.4 C 223kPa 4:39
136.3 C 220kPa 5:09
136.3 C 222kPa 5:39
136.1 C 222kPa 6:09
136.4 C 224kPa 6:39
136.4 C 223kPa 7:09
136.4 C 223kPa 7:09
VENTING 7:09
AIR DRYING 7:56
CYCLE COMPLETE 8:51

Digital Signature #
C7ED2D88A849ED77
