
Midmark M11 - v1.0.0
Total Cycles: 4996

Name: RITTER 1
Sterilizer ID: V2178198
10/15/2024 01:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 85.6 F 0.0
2:00 88.3 F 0.0
4:00 115.0 F 0.1
6:00 144.8 F 0.5
8:00 176.0 F 1.3
10:00 194.3 F 4.3
12:00 201.8 F 7.1
14:00 213.7 F 10.5
16:00 221.8 F 8.0
18:00 227.8 F 8.3
20:00 238.7 F 12.7
22:00 237.1 F 9.8
24:00 238.5 F 9.4
26:00 247.3 F 13.8
28:00 258.4 F 19.9
30:00 268.6 F 26.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.3
1:00 272.3 F 28.8
2:00 272.3 F 28.5
3:00 271.8 F 28.4
4:00 272.7 F 29.0

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.9 F 0.0
5:00 204.8 F 0.0
10:00 202.5 F 0.0
15:00 195.7 F 0.0
20:00 207.7 F 0.0
25:00 215.5 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 31:07
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:09:35
CYCLE COMPLETE

10/15/2024 02:26 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____