
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001328
9:01 31/10/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:05
PRESSURIZING 2:13
STERILIZING 3:17
136.0 C 220kPa 3:17
136.3 C 223kPa 3:47
136.3 C 223kPa 4:17
136.2 C 223kPa 4:47
136.4 C 224kPa 5:17
136.4 C 224kPa 5:47
136.1 C 222kPa 6:17
136.1 C 221kPa 6:47
136.1 C 221kPa 6:47
VENTING 6:47
AIR DRYING 7:37
CYCLE COMPLETE 12:31

Digital Signature #
D39F76C353AFB7A2
