

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3904

Name: RITTER 1  
Sterilizer ID: V2178198  
10/16/2023 08:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.2 F 0.0  
2:00 70.2 F 0.0  
4:00 75.6 F 0.0  
6:00 97.3 F 0.1  
8:00 131.5 F 0.3  
10:00 164.9 F 3.0  
12:00 182.4 F 6.3  
14:00 199.4 F 9.5  
16:00 214.0 F 13.0  
18:00 225.7 F 6.4  
20:00 235.3 F 10.4  
22:00 237.3 F 8.8  
24:00 242.4 F 12.0  
26:00 238.7 F 9.9  
28:00 251.3 F 15.9  
30:00 261.7 F 21.8  
32:00 271.5 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.5 F 29.0  
2:00 271.4 F 28.1  
3:00 272.5 F 29.1  
4:00 271.9 F 28.5  
  
Min 271.2 F 28.1  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 200.9 F 0.0  
5:00 220.1 F 0.0  
10:00 233.4 F 0.0  
15:00 229.9 F 0.0  
20:00 229.6 F 0.0  
25:00 229.0 F 0.0  
30:00 228.1 F 0.0  
  
Drying Complete

FILLING: 0:50  
HEATING: 32:41  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:10:56  
CYCLE COMPLETE

10/16/2023 09:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_