
Midmark M11 - v1.0.5
Total Cycles: 753

Name: RITTER2
Sterilizer ID: V2313228
02/02/2022 02:28 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.9 F 0.0
2:00 143.7 F 2.0
4:00 171.5 F 5.9
6:00 194.9 F 10.8
8:00 214.9 F 5.0
10:00 228.4 F 10.5
12:00 231.0 F 7.6
14:00 242.6 F 12.3
16:00 244.4 F 12.9
18:00 257.0 F 19.7
20:00 268.3 F 26.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.7 F 29.9
2:00 272.3 F 29.3
3:00 271.3 F 28.9
4:00 272.1 F 29.7

Min 271.2 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.3 F 0.0
5:00 217.6 F 0.0
10:00 217.0 F 0.0
15:00 214.8 F 0.0
20:00 227.7 F 0.0
25:00 226.7 F 0.0
30:00 217.0 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:59:33
CYCLE COMPLETE

02/02/2022 03:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____