
Midmark M11 - v1.0.5
Total Cycles: 76

Name: RITTER2
Sterilizer ID: V2313228
06/08/2021 09:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.0 F 0.0
2:00 179.8 F 3.9
4:00 199.4 F 10.0
6:00 214.2 F 5.3
8:00 231.8 F 12.1
10:00 236.0 F 10.9
12:00 239.2 F 10.5
14:00 254.1 F 18.0
16:00 266.9 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 29.4
1:00 271.8 F 29.5
2:00 271.8 F 29.5
3:00 271.7 F 29.3
4:00 272.2 F 29.4

Min 271.6 F 29.2
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.3 F 0.0
5:00 242.1 F 0.0
10:00 228.1 F 0.0
15:00 227.8 F 0.0
20:00 222.2 F 0.0
25:00 219.5 F 0.0
30:00 184.5 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 17:19
STERILIZING: 4:00
VENTING: 2:58
DRYING: 30:00
TOTAL CYCLE: 00:56:19
CYCLE COMPLETE

06/08/2021 10:40 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____