

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001881  
10:21 23/07/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:43  
PRESSURIZING 2:51  
STERILIZING 3:53  
136.0 C 220kPa 3:53  
136.3 C 223kPa 4:23  
136.4 C 225kPa 4:53  
136.2 C 222kPa 5:23  
136.4 C 225kPa 5:53  
136.4 C 225kPa 6:23  
136.3 C 222kPa 6:53  
136.2 C 222kPa 7:23  
136.2 C 222kPa 7:23  
VENTING 7:23  
AIR DRYING 8:13  
CYCLE COMPLETE 8:44  
-----

Digital Signature #  
A56308542709AA92  
-----