
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 006185
14:48 28/08/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:06
STERILIZING 4:15
136.5 C 225kPa 4:15
136.4 C 222kPa 4:45
136.4 C 223kPa 5:15
136.3 C 224kPa 5:45
136.4 C 222kPa 6:15
136.4 C 221kPa 6:45
136.4 C 225kPa 7:15
136.4 C 225kPa 7:45
136.4 C 225kPa 7:45
VENTING 7:45
AIR DRYING 8:35
CYCLE COMPLETE 8:50

Digital Signature #
200DE24AF969A8AD
