
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000247
14:04 29/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:29
PRESSURIZING 3:31
STERILIZING 5:02
136.0 C 230kPa 5:02
136.4 C 231kPa 5:32
136.3 C 232kPa 6:02
136.4 C 233kPa 6:32
136.5 C 233kPa 7:02
136.4 C 233kPa 7:32
136.4 C 233kPa 8:02
136.3 C 232kPa 8:32
136.3 C 232kPa 8:32
VENTING 8:33
AIR DRYING 9:34
CYCLE COMPLETE 16:35

Digital Signature #
8898F9344391A457
