

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.1uS / 0.7ppm  
CYCLE NUMBER 000309  
10:10 20/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:31  
STERILIZING 3:44  
136.0 C 228kPa 3:44  
136.1 C 231kPa 4:14  
136.3 C 232kPa 4:44  
136.4 C 233kPa 5:14  
136.2 C 231kPa 5:44  
136.3 C 231kPa 6:14  
136.3 C 230kPa 6:44  
136.3 C 233kPa 7:14  
136.3 C 233kPa 7:14  
VENTING 7:14  
AIR DRYING 8:09  
CYCLE COMPLETE 8:18  
-----

Digital Signature #  
8499F5886ACDB12A  
-----