
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.3uS / 0.8ppm
CYCLE NUMBER 000626
13:30 02/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CYCLE ABORTED 0:47
STOP BUTTON PRESSED
NOT STERILE
END TIME 0:47

Digital Signature #
298E03285F0B5F6D
