
Midmark M11 - v1.0.0
Total Cycles: 4371

Name: RITTER 1
Sterilizer ID: V2178198
03/08/2024 12:19 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.0 F 0.0
2:00 145.1 F 2.3
4:00 161.8 F 5.5
6:00 184.5 F 9.2
8:00 203.1 F 13.6
10:00 216.3 F 6.7
12:00 228.4 F 11.0
14:00 229.5 F 6.8
16:00 240.5 F 12.2
18:00 239.6 F 10.1
20:00 251.9 F 16.1
22:00 262.4 F 22.1
24:00 272.3 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.8 F 28.9
2:00 272.4 F 29.2
3:00 271.9 F 28.6
4:00 272.4 F 29.2

Min 271.4 F 28.5
Max 273.2 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.0 F 0.0
5:00 236.8 F 0.0
10:00 224.5 F 0.0
15:00 220.8 F 0.0
20:00 226.8 F 0.0
25:00 234.2 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 25:12
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:59
CYCLE COMPLETE

03/08/2024 01:22 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____