

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1556

Name: RITTER2  
Sterilizer ID: V2313228  
10/03/2022 03:07 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.6 F 0.0  
2:00 196.1 F 1.5  
4:00 214.2 F 6.7  
6:00 230.8 F 12.3  
8:00 234.2 F 9.3  
10:00 236.9 F 9.3  
12:00 243.1 F 11.6  
14:00 247.0 F 14.0  
16:00 259.7 F 21.1  
18:00 270.9 F 28.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 28.9  
1:00 272.8 F 29.6  
2:00 271.6 F 28.8  
3:00 272.8 F 29.5  
4:00 271.6 F 28.8  
  
Min 271.6 F 28.8  
Max 272.9 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.2 F 0.0  
5:00 225.9 F 0.0  
10:00 243.6 F 0.0  
15:00 236.1 F 0.0  
20:00 223.3 F 0.0  
25:00 218.3 F 0.0  
30:00 215.9 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 18:34  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:56  
CYCLE COMPLETE

10/03/2022 04:04 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_