

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1920

Name: RITTER2  
Sterilizer ID: V2313228  
01/11/2023 01:01 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 199.8 F 0.0  
2:00 195.3 F 4.3  
4:00 204.8 F 9.9  
6:00 221.8 F 6.3  
8:00 235.8 F 12.0  
10:00 238.1 F 10.3  
12:00 237.1 F 9.4  
14:00 252.3 F 16.9  
16:00 265.0 F 24.4

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.9  
1:00 272.2 F 29.5  
2:00 271.3 F 28.9  
3:00 272.1 F 29.4  
4:00 271.2 F 28.9  
  
Min 271.2 F 28.7  
Max 272.3 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.3 F 0.0  
5:00 231.4 F 0.0  
10:00 233.7 F 0.0  
15:00 228.4 F 0.0  
20:00 224.0 F 0.0  
25:00 234.5 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 17:34  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:56:55  
CYCLE COMPLETE

01/11/2023 01:57 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_