
Midmark M11 - v1.0.0
Total Cycles: 4610

Name: RITTER 1
Sterilizer ID: V2178198
06/05/2024 12:01 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 128.8 F 0.0
2:00 129.2 F 1.5
4:00 151.3 F 4.4
6:00 186.7 F 7.8
8:00 210.1 F 12.2
10:00 217.4 F 5.0
12:00 228.2 F 8.9
14:00 237.6 F 13.1
16:00 235.7 F 9.1
18:00 240.8 F 11.0
20:00 243.3 F 12.1
22:00 254.3 F 17.8
24:00 264.4 F 23.8
26:00 271.3 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.9
1:00 272.2 F 29.0
2:00 272.9 F 29.3
3:00 271.5 F 29.0
4:00 272.0 F 29.0

Min 271.3 F 28.8
Max 273.1 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 229.1 F 0.0
10:00 218.0 F 0.0
15:00 201.8 F 0.0
20:00 211.8 F 0.0
25:00 232.1 F 0.0
30:00 226.9 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:52
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:05:28
CYCLE COMPLETE

06/05/2024 01:06 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____