

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4233

Name: RITTER 1  
Sterilizer ID: V2178198  
01/25/2024 03:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.7 F 0.0  
2:00 101.7 F 0.8  
4:00 112.9 F 2.8  
6:00 141.7 F 5.6  
8:00 181.0 F 8.9  
10:00 196.5 F 13.4  
12:00 207.8 F 4.8  
14:00 217.1 F 8.5  
16:00 229.0 F 12.7  
18:00 232.2 F 8.5  
20:00 243.5 F 13.5  
22:00 241.9 F 11.4  
24:00 253.5 F 17.2  
26:00 263.8 F 23.3  
28:00 272.2 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.4  
1:00 272.6 F 29.1  
2:00 271.8 F 28.5  
3:00 271.6 F 28.7  
4:00 272.6 F 29.1

Min 271.2 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.7 F 0.0  
5:00 240.1 F 0.0  
10:00 228.1 F 0.0  
15:00 222.7 F 0.0  
20:00 227.0 F 0.0  
25:00 229.8 F 0.0  
30:00 225.0 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 28:14  
STERILIZING: 4:00  
VENTING: 3:33  
DRYING: 30:00  
TOTAL CYCLE: 01:06:49  
CYCLE COMPLETE

01/25/2024 04:46 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_