

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1025

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2022 09:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.8 F 0.0  
2:00 180.7 F 3.1  
4:00 188.0 F 7.6  
6:00 229.1 F 13.1  
8:00 227.5 F 8.4  
10:00 237.5 F 11.1  
12:00 242.4 F 12.4  
14:00 242.3 F 11.7  
16:00 255.5 F 18.7  
18:00 267.3 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.9  
1:00 272.2 F 29.1  
2:00 272.1 F 29.6  
3:00 271.4 F 28.8  
4:00 272.1 F 29.5  
  
Min 271.0 F 28.7  
Max 272.5 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.9 F 0.0  
5:00 237.5 F 0.0  
10:00 222.2 F 0.0  
15:00 214.3 F 0.0  
20:00 193.8 F 0.0  
25:00 191.5 F 0.0  
30:00 185.8 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:18  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:58:06  
CYCLE COMPLETE

04/25/2022 10:56 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_