

-----  
Midmark M11 - v1.0.5  
Total Cycles: 816

Name: RITTER2  
Sterilizer ID: V2313228  
02/17/2022 01:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 158.2 F 0.0  
2:00 156.4 F 2.9  
4:00 179.1 F 6.8  
6:00 200.5 F 11.6  
8:00 220.1 F 6.1  
10:00 234.0 F 11.4  
12:00 233.6 F 8.7  
14:00 239.0 F 10.5  
16:00 245.4 F 13.7  
18:00 258.0 F 20.3  
20:00 268.7 F 27.2

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.2  
1:00 272.4 F 29.6  
2:00 271.6 F 29.6  
3:00 271.9 F 29.2  
4:00 272.4 F 29.8

Min 271.1 F 29.0  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
5:00 226.4 F 0.0  
10:00 226.4 F 0.0  
15:00 223.1 F 0.0  
20:00 223.8 F 0.0  
25:00 230.9 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 1:35  
HEATING: 20:57  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:59:44  
CYCLE COMPLETE

02/17/2022 02:12 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02