

-----  
Midmark M11 - v1.0.5  
Total Cycles: 236

Name: RITTER2  
Sterilizer ID: V2313228  
07/22/2021 08:09 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 139.6 F 0.0  
2:00 142.4 F 1.9  
4:00 160.2 F 5.4  
6:00 186.6 F 9.9  
8:00 209.5 F 5.2  
10:00 222.6 F 8.3  
12:00 235.8 F 13.5  
14:00 238.7 F 10.8  
16:00 236.4 F 9.2  
18:00 250.2 F 15.8  
20:00 261.8 F 22.4  
22:00 272.3 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.9  
1:00 272.3 F 29.2  
2:00 272.5 F 29.7  
3:00 271.8 F 28.8  
4:00 272.5 F 29.6

Min 271.5 F 28.7  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 175.6 F 0.0  
5:00 226.4 F 0.0  
10:00 195.8 F 0.0  
15:00 192.3 F 0.0  
20:00 187.2 F 0.0  
25:00 183.1 F 0.0  
30:00 181.0 F 0.0  
Drying Complete

FILLING: 1:27  
HEATING: 22:26  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:00:58  
CYCLE COMPLETE

07/22/2021 09:10 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02