
Midmark M11 - v1.0.5
Total Cycles: 3394

Name: RITTER2
Sterilizer ID: V2313228
02/21/2024 09:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.8 F 0.0
2:00 105.6 F 0.0
4:00 110.8 F 0.5
6:00 137.5 F 2.9
8:00 167.8 F 6.4
10:00 191.7 F 10.4
12:00 210.3 F 6.7
14:00 217.2 F 8.0
16:00 228.2 F 12.4
18:00 232.8 F 9.0
20:00 238.2 F 10.5
22:00 245.1 F 13.3
24:00 256.4 F 19.5
26:00 267.0 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.9
1:00 272.4 F 29.9
2:00 272.5 F 29.9
3:00 272.4 F 29.9
4:00 272.3 F 30.0

Min 270.6 F 28.7
Max 273.3 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.7 F 0.0
5:00 207.0 F 0.0
10:00 186.6 F 0.0
15:00 183.3 F 0.0
20:00 180.0 F 0.0
25:00 177.7 F 0.0
30:00 176.2 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:27
STERILIZING: 4:00
VENTING: 3:28
DRYING: 30:00
TOTAL CYCLE: 01:05:52
CYCLE COMPLETE

02/21/2024 10:51 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____