
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001346
14:26 11/11/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:38
PRESSURIZING 2:45
STERILIZING 3:50
136.4 C 220kPa 3:50
136.4 C 226kPa 4:20
136.4 C 222kPa 4:50
136.4 C 225kPa 5:20
136.4 C 226kPa 5:50
136.4 C 226kPa 6:20
136.4 C 225kPa 6:50
136.4 C 226kPa 7:20
136.4 C 226kPa 7:20
VENTING 7:20
AIR DRYING 8:11
CYCLE COMPLETE 10:47

Digital Signature #
F6E60A2DD91F122A
