

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1002

Name: RITTER2  
Sterilizer ID: V2313228  
04/19/2022 08:21 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.1 F 0.0  
2:00 162.9 F 0.2  
4:00 169.9 F 0.7  
6:00 199.1 F 4.6  
8:00 212.9 F 9.3  
10:00 226.0 F 8.9  
12:00 233.4 F 10.7  
14:00 233.5 F 8.1  
16:00 240.9 F 11.0  
18:00 246.0 F 13.9  
20:00 258.3 F 20.6  
22:00 269.3 F 27.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.3  
1:00 272.4 F 29.5  
2:00 272.3 F 30.0  
3:00 271.7 F 29.2  
4:00 272.4 F 29.9

Min 271.5 F 29.1  
Max 272.6 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.2 F 0.0  
5:00 206.8 F 0.0  
10:00 209.9 F 0.0  
15:00 214.6 F 0.0  
20:00 219.0 F 0.0  
25:00 222.4 F 0.0  
30:00 226.0 F 0.0  
Drying Complete

FILLING: 1:13  
HEATING: 23:01  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:01:26  
CYCLE COMPLETE

04/19/2022 09:22 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_