
Midmark M11 - v1.0.0
Total Cycles: 3932

Name: RITTER 1
Sterilizer ID: V2178198
10/20/2023 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 171.5 F 3.3
4:00 213.5 F 7.5
6:00 208.9 F 13.1
8:00 224.1 F 7.3
10:00 237.0 F 12.4
12:00 236.6 F 9.5
14:00 237.3 F 9.1
16:00 248.0 F 14.4
18:00 260.1 F 21.0
20:00 271.1 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.4
1:00 272.2 F 28.8
2:00 271.4 F 28.6
3:00 272.4 F 28.8
4:00 271.3 F 28.5

Min 271.0 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 238.5 F 0.0
10:00 222.7 F 0.0
15:00 224.8 F 0.0
20:00 226.7 F 0.0
25:00 224.7 F 0.0
30:00 226.4 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 20:41
STERILIZING: 4:00
VENTING: 3:29
DRYING: 30:00
TOTAL CYCLE: 00:59:48
CYCLE COMPLETE

10/20/2023 04:37 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____