
Midmark M11 - v1.0.5
Total Cycles: 1936

Name: RITTER2
Sterilizer ID: V2313228
01/17/2023 10:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 179.7 F 2.8
4:00 181.1 F 6.9
6:00 202.9 F 12.0
8:00 217.2 F 6.9
10:00 233.5 F 12.5
12:00 235.7 F 10.0
14:00 236.3 F 8.8
16:00 248.7 F 15.3
18:00 260.8 F 22.1
20:00 271.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.2
1:00 272.0 F 29.9
2:00 272.2 F 29.6
3:00 271.2 F 29.3
4:00 272.6 F 30.0

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.1 F 0.0
5:00 229.5 F 0.0
10:00 229.5 F 0.0
15:00 228.8 F 0.0
20:00 228.9 F 0.0
25:00 228.6 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:27
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

01/17/2023 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____