
Midmark M11 - v1.0.5
Total Cycles: 786

Name: RITTER2
Sterilizer ID: V2313228
02/10/2022 05:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.5 F 0.0
2:00 169.9 F 3.0
4:00 189.7 F 7.2
6:00 211.5 F 12.7
8:00 222.8 F 7.7
10:00 239.9 F 13.4
12:00 239.2 F 11.3
14:00 238.4 F 10.2
16:00 252.7 F 17.1
18:00 264.6 F 24.3

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.3
1:00 272.1 F 29.4
2:00 272.5 F 29.9
3:00 271.6 F 29.4
4:00 272.4 F 29.6

Min 271.6 F 29.1
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.7 F 0.0
5:00 237.9 F 0.0
10:00 218.1 F 0.0
15:00 201.0 F 0.0
20:00 209.3 F 0.0
25:00 213.5 F 0.0
30:00 200.6 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 19:52
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

02/10/2022 06:54 PM
ACCEPTED: X
REJECTED:
Operator: U02