
Midmark M11 - v1.0.5
Total Cycles: 311

Name: RITTER2
Sterilizer ID: V2313228
08/25/2021 10:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 173.4 F 3.4
4:00 186.8 F 7.8
6:00 205.5 F 13.0
8:00 220.9 F 7.3
10:00 232.7 F 12.2
12:00 235.6 F 9.9
14:00 238.3 F 9.5
16:00 248.4 F 15.2
18:00 259.9 F 21.7
20:00 270.5 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.6 F 29.9
2:00 271.8 F 29.2
3:00 272.0 F 30.0
4:00 272.3 F 29.7

Min 271.2 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.2 F 0.0
5:00 226.9 F 0.0
10:00 224.2 F 0.0
15:00 221.2 F 0.0
20:00 213.8 F 0.0
25:00 205.5 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 2:00
HEATING: 20:49
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 01:00:08
CYCLE COMPLETE

08/25/2021 11:47 AM
ACCEPTED: X
REJECTED:
Operator: U02