

-----  
Midmark M11 - v1.0.5  
Total Cycles: 382

Name: RITTER2  
Sterilizer ID: V2313228  
09/22/2021 09:10 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 87.2 F 0.0  
2:00 88.5 F 0.4  
4:00 100.8 F 1.7  
6:00 132.5 F 4.1  
8:00 159.7 F 7.0  
10:00 174.9 F 10.1  
12:00 193.7 F 7.6  
14:00 210.6 F 5.8  
16:00 221.6 F 9.6  
18:00 230.4 F 11.1  
20:00 236.2 F 10.7  
22:00 235.2 F 8.6  
24:00 247.0 F 14.5  
26:00 257.5 F 20.2  
28:00 267.2 F 26.3

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.3  
1:00 272.5 F 29.7  
2:00 272.2 F 30.0  
3:00 271.6 F 29.2  
4:00 272.6 F 29.9  
  
Min 271.5 F 29.2  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.6 F 0.0  
5:00 214.1 F 0.0  
10:00 215.9 F 0.0  
15:00 192.1 F 0.0  
20:00 185.2 F 0.0  
25:00 190.3 F 0.0  
30:00 161.9 F 0.0  
Drying Complete

FILLING: 0:56  
HEATING: 29:30  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:07:44  
CYCLE COMPLETE

09/22/2021 10:17 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01