
Midmark M11 - v1.0.5
Total Cycles: 4263

Name: RITTER2
Sterilizer ID: V2313228
01/14/2025 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 90.3 F 0.0
2:00 90.4 F 0.0
4:00 96.4 F 0.0
6:00 128.2 F 0.1
8:00 163.7 F 0.3
10:00 186.3 F 0.5
12:00 200.5 F 2.4
14:00 219.4 F 5.6
16:00 229.9 F 9.1
18:00 238.9 F 12.9
20:00 235.5 F 9.4
22:00 240.5 F 11.2
24:00 243.9 F 12.6
26:00 241.9 F 11.5
28:00 253.4 F 17.3
30:00 263.6 F 23.4
32:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.7
1:00 272.6 F 29.4
2:00 271.6 F 28.6
3:00 272.3 F 29.5
4:00 271.9 F 29.0

Min 271.4 F 28.5
Max 272.6 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.8 F 0.0
5:00 230.1 F 0.0
10:00 221.7 F 0.0
15:00 217.3 F 0.0
20:00 218.5 F 0.0
25:00 229.2 F 0.0
30:00 227.1 F 0.0

Drying Complete

FILLING: 0:52
HEATING: 32:19
STERILIZING: 4:00
VENTING: 3:53
DRYING: 30:00
TOTAL CYCLE: 01:11:04
CYCLE COMPLETE

01/14/2025 04:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____