
Midmark M11 - v1.0.5
Total Cycles: 788

Name: RITTER2
Sterilizer ID: V2313228
02/11/2022 10:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	72.6 F	0.0
2:00	76.0 F	0.3
4:00	97.5 F	1.9
6:00	132.6 F	4.4
8:00	163.1 F	7.9
10:00	185.8 F	12.2
12:00	208.8 F	4.3
14:00	220.1 F	8.2

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 231F
CHAMBER PRESS 12.0PSI
** ITEMS NOT STERILE **

02/11/2022 11:07 AM