
Midmark M11 - v1.0.0
Total Cycles: 4454

Name: RITTER 1
Sterilizer ID: V2178198
04/12/2024 02:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 157.5 F 0.0
2:00 157.3 F 2.6
4:00 170.2 F 6.4
6:00 198.2 F 11.1
8:00 214.8 F 5.0
10:00 226.8 F 10.5
12:00 230.1 F 7.1
14:00 243.3 F 13.5
16:00 242.6 F 12.2
18:00 255.9 F 19.0
20:00 267.6 F 26.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.9
1:00 272.3 F 29.3
2:00 271.7 F 29.3
3:00 272.1 F 29.1
4:00 271.6 F 29.3

Min 271.2 F 28.7
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 213.5 F 0.0
10:00 213.6 F 0.0
15:00 214.5 F 0.0
20:00 216.6 F 0.0
25:00 218.5 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 21:18
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:59:52
CYCLE COMPLETE

04/12/2024 03:53 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____