

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1333

Name: RITTER2  
Sterilizer ID: V2313228  
07/21/2022 01:29 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 162.0 F 0.0  
2:00 166.4 F 0.3  
4:00 191.8 F 4.6  
6:00 212.3 F 10.2  
8:00 224.5 F 6.3  
10:00 237.5 F 11.9  
12:00 237.3 F 10.2  
14:00 236.2 F 9.1  
16:00 251.3 F 16.5  
18:00 263.5 F 23.6

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.0  
1:00 272.7 F 29.7  
2:00 271.6 F 28.8  
3:00 272.7 F 29.7  
4:00 271.7 F 28.9  
  
Min 271.6 F 28.8  
Max 272.9 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.5 F 0.0  
5:00 229.7 F 0.0  
10:00 228.8 F 0.0  
15:00 229.5 F 0.0  
20:00 225.3 F 0.0  
25:00 228.6 F 0.0  
30:00 228.7 F 0.0  
Drying Complete

FILLING: 1:50  
HEATING: 19:54  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:49  
CYCLE COMPLETE

07/21/2022 02:27 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_