
Midmark M11 - v1.0.5
Total Cycles: 977

Name: RITTER2
Sterilizer ID: V2313228
04/12/2022 08:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 84.5 F 0.0
4:00 84.7 F 0.0
6:00 104.0 F 0.0
8:00 152.8 F 0.2
10:00 202.1 F 0.4
12:00 214.5 F 2.1
14:00 223.6 F 5.3
16:00 230.9 F 8.6
18:00 238.1 F 12.5
20:00 234.9 F 9.1
22:00 240.8 F 11.0
24:00 244.9 F 13.1
26:00 244.0 F 12.4
28:00 256.4 F 18.9
30:00 267.3 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 272.5 F 29.1
2:00 272.0 F 29.0
3:00 272.4 F 28.9
4:00 272.1 F 29.1

Min 271.7 F 28.5
Max 272.8 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.9 F 0.0
5:00 205.9 F 0.0
10:00 171.0 F 0.0
15:00 168.8 F 0.0
20:00 163.2 F 0.0
25:00 159.9 F 0.0
30:00 156.5 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 31:27
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

04/12/2022 09:55 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____