
Midmark M11 - v1.0.5
Total Cycles: 2347

Name: RITTER2
Sterilizer ID: V2313228
05/10/2023 10:16 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.2 F 0.0
2:00 185.2 F 3.2
4:00 192.9 F 7.1
6:00 216.0 F 12.1
8:00 225.7 F 7.1
10:00 234.9 F 12.4
12:00 234.1 F 9.6
14:00 236.6 F 9.2
16:00 246.9 F 14.3
18:00 258.8 F 20.9
20:00 269.5 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.1 F 29.3
2:00 272.2 F 29.8
3:00 271.6 F 29.4
4:00 272.1 F 29.4

Min 271.4 F 29.1
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.9 F 0.0
5:00 231.6 F 0.0
10:00 225.4 F 0.0
15:00 220.0 F 0.0
20:00 218.6 F 0.0
25:00 226.2 F 0.0
30:00 231.2 F 0.0
Drying Complete

FILLING: 1:27
HEATING: 20:55
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:59:34
CYCLE COMPLETE

05/10/2023 11:15 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____