

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1542

Name: RITTER2  
Sterilizer ID: V2313228  
09/27/2022 10:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
2:00 177.4 F 3.4  
4:00 195.5 F 8.7  
6:00 223.5 F 7.7  
8:00 234.1 F 10.7  
10:00 233.0 F 8.6  
12:00 239.0 F 10.1  
14:00 248.3 F 15.0  
16:00 261.5 F 22.4  
18:00 272.7 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.0 F 29.1  
1:00 272.5 F 29.8  
2:00 271.4 F 29.0  
3:00 272.1 F 29.9  
4:00 272.3 F 29.6  
  
Min 271.0 F 28.9  
Max 272.8 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.8 F 0.0  
5:00 229.1 F 0.0  
10:00 213.7 F 0.0  
15:00 219.8 F 0.0  
20:00 223.2 F 0.0  
25:00 219.6 F 0.0  
30:00 210.0 F 0.0  
Drying Complete

FILLING: 3:01  
HEATING: 18:21  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:36  
CYCLE COMPLETE

09/27/2022 11:04 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_