

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 004271  
13:21 30/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:44  
PRESSURIZING 1:49  
STERILIZING 2:46  
136.4 C 221kPa 2:46  
136.4 C 221kPa 3:16  
136.4 C 224kPa 3:46  
136.2 C 223kPa 4:16  
136.4 C 223kPa 4:46  
136.4 C 225kPa 5:16  
136.4 C 225kPa 5:46  
136.4 C 224kPa 6:16  
136.4 C 224kPa 6:16  
VENTING 6:16  
AIR DRYING 7:10  
CYCLE COMPLETE 7:13

-----  
Digital Signature #  
BA9DF3A62E4534F3  
-----