
Midmark M11 - v1.0.5
Total Cycles: 815

Name: RITTER2
Sterilizer ID: V2313228
02/17/2022 11:44 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 175.8 F 3.5
4:00 187.7 F 8.4
6:00 210.6 F 10.0
8:00 226.3 F 9.5
10:00 234.3 F 9.0
12:00 242.6 F 12.8
14:00 243.6 F 13.1
16:00 257.0 F 20.1
18:00 268.6 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.6
1:00 272.7 F 30.2
2:00 271.7 F 29.4
3:00 272.6 F 30.4
4:00 272.4 F 30.0

Min 271.5 F 29.4
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.2 F 0.0
5:00 217.6 F 0.0
10:00 229.4 F 0.0
15:00 229.4 F 0.0
20:00 225.3 F 0.0
25:00 228.2 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 2:01
HEATING: 19:10
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

02/17/2022 01:09 PM
ACCEPTED: X
REJECTED:
Operator: U02