

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2763

Name: RITTER 1  
Sterilizer ID: V2178198  
01/11/2023 09:38 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
2:00 184.6 F 0.0  
4:00 190.0 F 1.5  
6:00 209.4 F 6.3  
8:00 221.1 F 11.5  
10:00 227.7 F 7.0  
12:00 240.5 F 12.5  
14:00 239.0 F 10.2  
16:00 237.7 F 8.2  
18:00 250.6 F 15.3  
20:00 262.4 F 22.1  
22:00 272.6 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.1  
1:00 272.6 F 28.8  
2:00 271.8 F 28.0  
3:00 272.2 F 28.8  
4:00 272.3 F 28.5

Min 271.4 F 27.9  
Max 272.6 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.9 F 0.0  
5:00 224.2 F 0.0  
10:00 232.0 F 0.0  
15:00 230.0 F 0.0  
20:00 226.7 F 0.0  
25:00 224.6 F 0.0  
30:00 224.6 F 0.0  
Drying Complete

FILLING: 0:48  
HEATING: 22:20  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:00:17  
CYCLE COMPLETE

01/11/2023 10:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_