
Midmark M11 - v1.0.0
Total Cycles: 1205

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2021 09:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 189.0 F 3.0
4:00 194.6 F 7.4
6:00 214.8 F 13.1
8:00 227.2 F 7.9
10:00 241.4 F 13.4
12:00 239.7 F 10.9
14:00 237.5 F 9.3
16:00 251.5 F 16.1
18:00 263.3 F 22.9
20:00 271.7 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.8
2:00 271.8 F 28.5
3:00 272.6 F 28.8
4:00 271.6 F 28.3

Min 271.5 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 214.3 F 0.0
5:00 225.5 F 0.0
10:00 196.4 F 0.0
15:00 189.6 F 0.0
20:00 184.4 F 0.0
25:00 183.6 F 0.0
30:00 181.5 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 20:04
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:58:51
CYCLE COMPLETE

09/27/2021 10:09 AM
ACCEPTED: X
REJECTED:
Operator: U01