
Midmark M11 - v1.0.0
Total Cycles: 2697

Name: RITTER 1
Sterilizer ID: V2178198
12/20/2022 10:26 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.1 F 0.0
2:00 166.8 F 0.2
4:00 179.9 F 2.4
6:00 205.0 F 7.0
8:00 222.0 F 12.8
10:00 227.1 F 7.8
12:00 240.3 F 13.4
14:00 240.4 F 11.0
16:00 238.4 F 9.4
18:00 252.9 F 16.5
20:00 265.0 F 23.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.5 F 28.7
2:00 271.5 F 28.2
3:00 272.5 F 28.8
4:00 271.5 F 28.0

Min 271.4 F 28.0
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.6 F 0.0
5:00 210.1 F 0.0
10:00 231.9 F 0.0
15:00 229.2 F 0.0
20:00 230.5 F 0.0
25:00 225.2 F 0.0
30:00 219.8 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:48
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:50
CYCLE COMPLETE

12/20/2022 11:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____