
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000996
11:37 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:37
STERILIZING 3:40
136.4 C 233kPa 3:40
136.3 C 232kPa 4:10
136.4 C 233kPa 4:40
136.4 C 233kPa 5:10
136.1 C 232kPa 5:40
136.4 C 233kPa 6:10
136.3 C 233kPa 6:40
136.1 C 232kPa 7:10
136.1 C 232kPa 7:10
VENTING 7:10
AIR DRYING 8:00
CYCLE COMPLETE 17:46

Digital Signature #
3C1AB1E829526CDE
