
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.7uS / 1.0ppm
CYCLE NUMBER 000807
15:24 09/11/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:09
STERILIZING 2:57
136.0 C 228kPa 2:57
136.2 C 230kPa 3:27
136.4 C 231kPa 3:57
135.7 C 227kPa 4:27
136.2 C 231kPa 4:57
136.2 C 230kPa 5:27
136.4 C 233kPa 5:57
136.4 C 230kPa 6:27
136.4 C 230kPa 6:27
VENTING 6:27
AIR DRYING 7:14
CYCLE COMPLETE 22:14

Digital Signature #
ADD484A87CC7B792
