

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 004223  
8:45 24/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:55  
STERILIZING 3:02  
136.5 C 222kPa 3:02  
136.3 C 222kPa 3:32  
136.0 C 219kPa 4:02  
136.4 C 222kPa 4:32  
136.1 C 221kPa 5:02  
136.3 C 222kPa 5:32  
136.1 C 217kPa 6:02  
136.2 C 219kPa 6:32  
136.2 C 219kPa 6:32  
VENTING 6:32  
AIR DRYING 7:24  
CYCLE COMPLETE 17:57

-----  
Digital Signature #  
B033D077D9ED41E0  
-----