
Midmark M11 - v1.0.0
Total Cycles: 1855

Name: RITTER 1
Sterilizer ID: V2178198
04/21/2022 12:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.7 F 0.0
2:00 190.9 F 4.7
4:00 218.4 F 10.7
6:00 225.5 F 6.1
8:00 239.6 F 12.1
10:00 238.6 F 10.0
12:00 236.6 F 8.5
14:00 251.8 F 16.0
16:00 264.2 F 23.1

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.5
2:00 271.8 F 28.3
3:00 272.4 F 28.5
4:00 271.6 F 28.1

Min 271.5 F 27.9
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 214.8 F 0.0
10:00 218.0 F 0.0
15:00 221.2 F 0.0
20:00 218.5 F 0.0
25:00 229.5 F 0.0
30:00 221.6 F 0.0
Drying Complete

FILLING: 2:27
HEATING: 17:54
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:57:31
CYCLE COMPLETE

04/21/2022 12:57 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____